

# **Research Use Only**

Maine Molecular Quality Controls, Inc. 23 Mill Brook Road, Saco, Maine 04072 USA Phone: 207-885-1072, FAX: 207-885-1079 Web: www.mmqci.com, Email: info@mmqci.com

## Birlinn<sup>™</sup> BCR-ABL1 p210 IS Control Panel C240

### **INTENDED USE:**

The Birlinn<sup>™</sup> BCR-ABL1 p210 IS Control Panel C240 is intended for use as an assayed external quality control to monitor the performance of in vitro laboratory nucleic acid detection and quantification of BCR-ABL1 (p210) fusion gene using molecular test systems. The Birlinn BCR-ABL1 p210 IS Control Panel C240, with assigned values traceable to World Health Organization (WHO), allows monitoring of assay performance reporting on the international scale.

The Philadelphia chromosome, a translocation between the ABL1 gene on chromosome 9 and the BCR gene on chromosome 22, designated as t(9;22), generates the fusion gene BCR-ABL1 which is present in most chronic myelogenous leukemia patients. Quantitative monitoring of BCR-ABL1 transcripts in patient blood is an important tool for measuring response to therapy. In 2009, the WHO developed a panel of four BCR-ABL1 primary standards to establish an international scale (IS), a standardized method for reporting assay results as a ratio of fusion transcripts to control gene transcripts (%IS), useful to the harmonization of patient care across laboratories worldwide.<sup>1,2</sup> The %IS can also be expressed as molecular response (MR), the log reduction from a standardized baseline of 100% on the IS. The Birlinn BCR-ABL1 p210 IS Control Panel C240 kit is traceable to the WHO International Genetic Reference Panel for Quantitation of BCR-ABL1 Translocation (WHO Reference Panel), NIBSC code 09/138.

### **PRODUCT SUMMARY and PRINCIPLE:**

The Birlinn BCR-ABL1 p210 IS Control Panel C240 consists of three components. Each component contains an increasing concentration of BCR-ABL1 (e14a2/b3a2) RNA transcript mixed with a fixed concentration of ABL1 RNA transcript to produce three levels 0.01%, 0.1%, and 1%. The %IS values provided are traceable to WHO Reference Panel, NIBSC code 09/138, and are assigned to each lot of the Birlinn BCR-ABL1 p210 IS Control Panel C240 according to NIBSC Instructions for Use.3,4

Quality controls can be used for routine monitoring of test systems, validation, verification, proficiency assessment, and training procedures. Consistent use of quality controls assists the laboratory in identifying shifts, trends, and increased frequency of random errors caused by variations in the test system, such as failing reagents or malfunctioning equipment. Early investigation can prevent failed assay runs

#### Validation and Value Assignment

MMQCI manufactured 3 lots of Birlinn BCR-ABL1 p210 IS Control Panel C240 and tested them alongside the WHO Reference Panel, using one reagent lot of the REALQUALITY RQ-BCR-ABL p210 One-Step assay. Linear Regression was applied and a lot-specific Correction Factor (CF) was calculated to assign WHOtraceable %IS values to each level of Birlinn BCR-ABL1 p210 IS Control Panel C240 for all 3 lots according to NIBSC Instructions for Use.3,4

#### **COMPOSITION:**

The Birlinn BCR-ABL1 p210 IS Control Panel C240 is comprised of 6 tubes, 2 tubes of each %IS level. The tubes contain 20µL of synthetic BCR-ABL1 RNA transcript and synthetic ABL1 control gene RNA transcript provided in a ready-touse solution. Each of the components can be directly used in a reverse transcription reaction, followed by amplification of cDNA for detection of BCR-ABL1/ABL1 using the laboratory's system. The Birlinn BCR-ABL1 p210 IS Control Panel C240 materials are provided in a background of total RNA at a concentration of 400ng/µL and may be diluted as a patient sample to provide the recommended sample input for the assay to be used.

## **STORAGE and STABILITY:**

The Birlinn BCR-ABL1 p210 IS Control Panel C240 should be stored at -25°C to -15°C and is stable through the expiration date printed on the kit label when stored frozen. Do not freeze/thaw more than 5 times. Discard after use according to your local and federal regulations.

- This product is intended for in vitro analytical testing and is provided for Research Use Only, not for use in diagnostic procedures.
- This product is clear in appearance.
- Do not freeze/thaw more than 5 times.
- This product does not contain any biological material of human or animal origin. Universal Precautions are NOT required when handling this product.
- The Birlinn BCR-ABL1 p210 IS Control Panel C240 cannot be cloned, sold, or transferred without the explicit written consent of MMQCI.

#### **RECOMMENDED INSTRUCTION FOR USE:**

- Allow the Birlinn BCR-ABL1 p210 IS Control Panel C240 component to be tested to come <u>completely</u> to room temperature (18°C-25°C), approximately 10-15 minutes.
- Immediately before pipetting, thoroughly mix the controls by flicking the 2. tube 20x followed by 2 pulse vortexes, 2-3 seconds each at maximum speed. Quick spin tube to remove droplets from cap.
- 3. Carefully pipette each component according to the assay Manufacturer's instructions for patient samples. If a dilution is required for sample concentrations of 400ng/µL, dilute the Birlinn BCR-ABL1 p210 IS Control Panel C240 components in the same manner. Note: It is recommended to use a minimum of 5µL of the Birlinn BCR-ABL1 p210 IS Control Panel C240 material for optimal performance.

Dilution Protocol Example:

- In a labeled 1.5mL LoBind tube, pipette  $5\mu L$  of well-mixed Birlinn 1. BCR-ABL1 p210 IS Control Panel C240 component into 45µL of molecular grade H2O to make a 1:10 dilution.
- 2. Mix the dilution by 20x inversions, followed by 2 pulse vortexes, 3-5 seconds. Quick spin to remove all liquid from cap prior to use.
- 4. Continue with the assay procedure according to manufacturer's instructions.
- 5. Discard after use according to local and federal regulations, or store remaining material at -25°C to -15°C.

## **EXPECTED VALUES:**

Locate the appropriate WHO-traceable % IS values assigned to your lot of Birlinn BCR-ABL p210 IS Control Panel C240 on the lot specific Certificate of Analysis. It is important to notice that the WHO-traceable values were assigned by testing with one lot of REALQUALITY RQ-BCR-ABL p210 One-Step assay. The reported % values of C240 may vary among laboratories, reagent lots, operators and test systems. Each laboratory should establish their own % ranges.

### **ORDERING INFORMATION:**

Birlinn BCR-ABL1 p210 IS Control Panel C240

Part Number: C240

Kit Contains: 6 tubes x 20µL;

2 tubes each of 0.01% IS, 0.1% IS, and 1% IS

References

<sup>1</sup>Branford S et al. Desirable performance characteristics for BCR-ABL measurement on an international reporting scale to allow consistent interpretation of individual patient response and comparison of response rates between clinical trials. Blood 2008, 112:3330-38. <sup>2</sup>White HE et al. Establishment of the first World Health Organization International Genetic

Reference Panel for quantitation of BCR-ABL mRNA. Blood 2010, 116:e111-117. <sup>3</sup>Cross et al. Development and evaluation of a secondary reference panel for BCR\_ABL1

quantification on the International Scale. Leukemia 2016 30, 1844-1852.

<sup>4</sup>Ist WHO International Genetic Reference Panel for Quantitation of BCR-ABL Translocation, NIBSC code: 09/138 Instructions for use (Version 9.0, Dated 22/10/2020).